Air Ion Counter AIC2



Operation Overview

During operation, air is drawn in at the top, is measured, and exits at the bottom. Do not block the air intake and exit openings. When operating, the display shows the ion count present in the air, updating once per second. The 20-foot long ground wire should be used to connect the meter to earth ground for most types of measurements, especially if you are measuring high ion levels.

At power on or when changing polarity, the meter stabilizes for about four seconds, after which ion measurements begin. For accurate readings, avoid measuring near charged plastic such as synthetic fabric (clothing). Excess charges will reduce both positive and negative ion counts in the vicinity.

In a typical environment, ions do not mix well. There can be clusters of negative ions in one area with almost no ions just a few inches away. Thus, it is normal for the ion level to fluctuate irregularly as air is drawn into the meter. To provide a more stable reading the meter imposes five second averaging. An alternative average of 14 seconds can be selected.

lon peak levels are continually saved in the meter. By Pressing the "Peak Mode" the saved maximum ion level that occurred (since the last time "Reset Peak" was depressed) is displayed.

"Record" initializes automated recording for later retrieval via the USB port. The AlphaApp PC application provides for retrieval and management of the meter data. The application may be downloaded at: https://www.alphalabinc.com/alphaapp/

"Scale" (dual range moderl only) switches the meters range between 10 to 2 million ion cm³ and 100 to 200 million ions cm³.

AIC2 Operation Details

Power Up / Operation

On power up the fan will come on, and the display will indicate "Stabilizing" for about four seconds, at which point ion measurements will begin. To change the desired ion polarity, press "Polarity." Note that when the polarity is changed, the meter restabilizes.

"Scale" selects the alternative 200 million range on the dual scale model. On the standard (single range) AIC2 the scale is fixed at 2 million. The 200 million scale allows for the measurement of extremely high ion levels, such as the output of an ion generator. A display of 1---.-- indicates over range.

Selected Polarity Temperature

Averaging

The factory default is 5 second averaging. The alternative averaging rate is 14 seconds. The 14 second averaging greatly reduces the ion level fluctuation caused by uneven mixing of ion levels in the air. However, the longer averaging results in a corresponding delay in the meter's response time.

To view the current averaging setting, hold down "Scale". Then (while continuing to hold "Scale" down), depress "Record". To change the average, follow the same procedure but hold both "Scale" and "Record" down for 10 seconds or until alternative average time appears.

Note that although the display of the average will disappear five seconds after depressing any other function, the selected average will continued to be employed until it is changed.





Default Range, 2 million ions cm³

Peak Mode

Pressing "Peak Mode" displays the real time ion levels (top displayed value) as well as the highest ion level measured by the meter since the last time the peak was reset. While in this mode, momentarily pressing "Peak Reset" will reset the peak value. When the peak mode is off, the meter continues to capture the highest ion level measured. When the polarity is changed, the peak value is reset.

Meter Calibration / Re-zeroing

Calibration establishes the zero point when no ions are present. Calibration is not necessary each time the meter is used. If there has been a major temperature variation or if the meter has not been used for a long period, the meters zero point should be verified. To do so, depress "Standby". In standby, the meter's ion sensor remains active, but the fan is turned off. After

stabilizing, the ion count should gradually drop to 0.00 +/- 0.06. Residual air ions and ambient noise will cause the reading to fluctuate by about two to six (least significant digit) counts. If the meter does not settle near 0.00, the meter needs to be recalibrated. For instance if the meter settles at a reading of -0.15, the meter's zero point has an offset of -0.15 which would be eliminated if recalibrated. It is advisable to be in the normal 5 second averaging before going into standby. In 5 second averaging the ion count will settle to zero in 10 to 20 seconds. In 14 second averaging settling takes more time. To come out of standby mode, depress "Standby" or any other function key.

To re-establish the meter's zero point, press "Calibrate". During the calibration process the zero reference point for both scales is established. During the process, a self-test is also conducted. Calibration typically takes 50 to 60 seconds, during which a status bar indicates the calibration progress. It is normal for the status bar to increment in steps at 10 to 15 second intervals.

Calibration and self test errors: If the zero point cannot be determined, a calibration timeout message will appear. This is caused by high wind, close proximity to a strong static or electrical field, or extremely high levels of ions. If this error occurs, cover the intake or place the meter in a different position/location. Other possible errors include: "Excessive Offset" or "Low Sensitivity" which is an indication of sensor contamination or moisture condensation. These errors can usually be resolved by blowing compressed air through the air intake, or in the case of condensation, allowing the meter to dry out.



23.2°C







Power Options and Battery Level

Operation time with alkaline batteries is typically 16 hours. The meter may be operated with a 6 or 9VDC, 200mA regulated or unregulated power adapter with a 2.1mm barrel plug, (positive center). Note that if the battery level drops below the level required for the internal processor, the display will be blank; however, the backlight of the display may still be illuminated. If this occurs, replace the batteries.



Battery Level: Each missing bar represents a drop of approximately 20% of battery's usable capacity.

Data Recording / USB

With the data recording function, measurements may be recorded for later upload to a PC/Laptop. When "Record" is pressed, the current displayed measurement is saved into non-volatile memory and a record number is momentary displayed. Up to 9,999 records or record sets may be saved. The software application AlphaApp provides the user with the ability to retrieve recorded data and delete saved data. The AlphaApp application also allows the option of configuring the meter to record a stream of measurements at a specified interval. When configured to record a stream of measurements, pressing "Record" turns on the record function indicated by the display of the record set number. In this mode measurements are continually saved at the configured interval until "Record" is pressed a second time, at which time the recording stops and the set number is no longer displayed. When attached to a PC/Laptop, the AlphaApp application also allows for the real-time streaming of the meter's measurements.



Record or Set #

AlphaApps USB Utility Download at: https://www.alphalabinc.com/alphaapp/

- Real time ion level data and plotting
- Upload/view/plot prerecorded data
- Selectable Ion measurement/record intervals; 0.5, 1.0, 2.5, 5, 10, 15, 30, or 120 seconds
- 96 hours of nonvolatile storage in the AIC2 at the 1 second measurement interval
- Ion level verses time graphs with optional selectable automatic scaling and time compression
- Plotted graphs and data tables may be saved for later retrieval
- Table data (which is comma-delimited) can be pasted into a spreadsheet







SPECIFICATIONS	
Ion Polarity	Positive or Negative lons - selectable
Ion Range Standard Model	2 Million cm ³
Dual Range Model	2 Million or 200 Million lons cm ³ - selectable
Ion Resolution	10 lons/cm ³ (2M Scale), 100 lons/cm ³ (200M Scale)
Ion Accuracy	+/- 20%, repeatability 5%
Averaging	5 or 14 seconds - selectable
Peak Value	Continually saved
	Peak value display - selectable
	Peak value reset - selectable
Real Time Data Output	Micro USB Port
Data Recording	On command or automatic at selectable intervals
	Up to 9,999 records - non-volatile memory
PC Application	Real time data plotting and record retrieval
AlphaApp	
Temperature	Displayed in Celsius, 1% accuracy
Calibration	Zero reference verification and Self Test- selectable
Environmental	-1° C to 43° C, 0-90% RH non-condensing
	Wind ,< 15 km/hr (9 mph)
Power - Internal	Battery - 4 AA Alkaline (~16 hour life)
	Optional -4 AA NiMh (~18 hour life)
	Consumption 75mA, 17mA in standby mode
	Battery condition continually displayed
Power - External	6 to 9VDC regulated or unregulated
	Power Jack, 2.1mm positive center
	9VDC regulated, 200mA Adapter included
Size	7.63 x 4 x 1.75 inches (194 x 102 x 45 mm)
Desk Stand	Retractable - Integrated into the bottom of the meter
Display	Graphic 64x128 pixels, Backlit, 1.3x2.5 inches (35x64mm)
Weight	20 oz (567 g)



Removable Shield on the Dual Range Model

Remove when using the 200 million ions /__ $x10^6$ scale.

This shield minimizes the interference of electric fields when measuring low levels of ions . When measuring high ion levels (200 million ions $(_x10^6)$ scale, it is important that the shield be removed, otherwise high concentrations of ions can be significantly dissipated before reaching the internal ion sensor. In the 200 million ion scale, the meter is not sensitive to electric fields.





Optional Carrying Case 10.6 x 11.8 x 2.7 inches (27 x 30 x 6.8cm)

Warranty: The warranty period is one year from date of delivery.

Designed and manufactured in the USA by AlphaLab Inc. 3005 South 300 West, Salt Lake City UT 84115 www.alphalabinc.com Tel (801)487-9492 Fax (801)487-3877

Manual V2.4 June 2019